
	Document No.	DC-102	Issue No.	1.0	Page	1 of 1
	Prepared By	John Adcock			Date	30 April 2010
DECLARATION OF CONFORMITY European Communities Council Directives 93/42/EEC and 2007/42/EC Concerning Medical Devices PRESSURE RELIEVING PATIENT CHAIRS AND ASSOCIATED CUSHIONS						

The undersigned declares that the Products listed below meet the provisions of the Council Directives which apply and the CE Mark may therefore be affixed. The Products are registered with the UK Competent Authority. All supporting documentation is retained at the premises of the manufacturer.

Product Description	Pressure Relieving Patient Chairs and Associated Cushions
Manufacturer	The Kirton Healthcare Group Ltd, 23 Rookwood Way, Haverhill, Suffolk, CB9 8PB, United Kingdom
Intended Use	Seating intervention for preventing pressure ulcers for people at very high risk, and assist in healing existing pressure ulcers up to Grade 2
MDD Classification	Class I (one) - Annex IX, Rule 1
Notified Body	Not applicable
EU Authorised Representative	Not applicable
Assessment route	Self certification by Annex VII; EC Declaration of Conformity and Article 14; Registration of persons responsible for placing devices on the market.
Applicable product standards and guidelines	BS EN ISO 14971:2009 Medical devices -- Application of risk management to medical devices EN 980:2008 Symbols for use in the labelling of medical devices EN 1041:2008 Information supplied by the manufacturer of medical devices BS EN 1021-1:2006 Furniture. Assessment of the ignitability of upholstered furniture. Ignition source smouldering cigarette. BS EN 1021-2:2006 Furniture, Upholstered furniture, Fire tests, Ignitability, Test equipment, Reports, Testing conditions BS 5852:2006 Methods of test for assessment of the ignitability of upholstered seating by smouldering and flaming ignition sources

Signed  Date 12/5/10

Name (printed) ANDREW HARDMAN

Position MANAGING DIRECTOR

Date	Issue No.	Amendment Description
30 April 2010	1.0	New document