

Test Certificate

This certificate confirms that the

Cushion components – Ref: Fabric: Dartex PU coated fabric; Interliner: Endogard FR fabric; Foam: R400H PU foam with 12” x 13” Intelli-gel insert

Supplied by

Kirton Healthcare Group Limited

has been tested at FIRA International Limited and successfully satisfied the requirements from:

BS EN 1021-1: 2006 (cigarette), BS EN 1021-2: 2006 (match flame equivalent) & BS 5852: 2006 Clause 11 Ignition Source 5.

Report reference: TFFLF27872

SIGNATURE

Stephen Cotton

POSITION

Technical Specialist - Flammability

CERTIFICATE DATE

14 May 2009

For and on behalf of FIRA INTERNATIONAL LIMITED

This certificate only relates to the sample(s) supplied and tested at the time. Re-testing at intervals is recommended and should be subject to agreement between the supplier and the purchaser.