

Test Certificate



This certificate confirms that the

PATIENT CHAIR

Supplied by

Kirton Healthcare Group Limited

has been tested at FIRA International Limited and successfully satisfied the applicable test requirements from:

BS EN 15373:2007 Test Level 3

BS EN 1022:2005

According to the Standard the Patient Chair is therefore considered to be structurally suitable and sufficiently stable for severe contract use.

Report reference: TSSEF33828 refers

SIGNATURE

A P Cayless

POSITION

Technical Specialist - Seating

CERTIFICATE DATE

18 August 2010

For and on behalf of FIRA INTERNATIONAL LIMITED



Expertise • Independence • Integrity

This certificate only relates to the sample(s) supplied and tested at the time. Re-testing at intervals is recommended and should be subject to agreement between the supplier and the purchaser.